

26, Man Singh Road, Jaisalmer House,  
New Delhi, 29.7.2010

To, <sup>11/04</sup> <sup>P.B.No-5/07</sup> <sup>11/02 5/08 5/07</sup>  
The DGs: Assam Rifles/BSF/CISF/CRPF/ITBP/NSG/SSB/BPR&D

Subject:- Specifications of various Medical/Hospital Equipments, Laboratory equipments and Laundry equipments -regarding


Sir,

The Specifications for the following Medical/Hospital Equipments, Laboratory equipments and Laundry equipments have been accepted and approved by the Competent Authority in MHA and the same are enclosed for information and Record:-

1. Condom Vending Machine- Annex-I
2. Flowcytometer- Annex-II
3. **Laboratory Equipments**
  - a) Coagulation Analyzer- Annex-III
  - b) Glycosylated Haemoglobin/Micro Albumin/CRP/D-Dimmer Analyzer-IV
  - c) Urine Analyzer – Annex-V
  - d) Semen Quality Analyzer- Annex-VI
  - e) Auto Pipette System- Annex-VII
  - f) Anaesthesia Work Station- Annex-VIII
- 4) **Laundry equipments**- Annex-IX
  - a) Washing Machine
  - b) Drier Washing Machine- 30 Kg (Moderate size)
  - c) Heavy Electric Press(Hand Operated)
  - d) Hydro Extractor
  - e) Table for pressing hospital linen
  - g) Sluice machine
  - h) Laundry Trolley
  - i) Bulk sterilizer electric for sterilizing linen and mattresses

2. Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Specifications.

Yours faithfully,



(S.B.Nanda)

Under Secretary to the Govt. of India

**Encls: as above**

Copy to:-DD(Procurement),MHA

Copy for information to:-PS to JS(PM),MHA

TECHNICAL SPECIFICATIONS OF COAGULATION ANALYZER

1. It should be Semi-Automated Coagulation single channel Analyzer.
2. Principal of Equipment should be operated on Mechanical ball method
3. It should report results in Second, ratio and INR
4. It should have the facility for attachment with Auto Starter Pipette
5. It should have the facility to maintain the Temp, atleast 37 degree centigrade
6. It Should Be Portable, Compact & User Friendly
7. It should have the facility to connect External Serial Printer.
8. It Should Contain At Least Single Measuring Chamber, At Least 4 Cuvette Incubation Chambers & One Reagent Incubation Chamber
9. Guarantee/Warranty/AMC must be ensured.

Presiding officer	<u>Dr. N.K. Bhambri</u> IG(Med)/Medical Supdt, of Referral Hospital
Member from BSF	<u>Dr. Rohit Kumar, CMO</u>
Member from CRPF	<u>Dr. Bipul Kumar, MO</u>
Member from ITBP	<u>Dr. Sandhya Thapliyal, SMO</u>
Member from SSB	<u>Dr. Brijesh Kumar, CMO (SG)</u>
Co-op Member	<u>Dr. B.S. Mishra</u> IG (Medical) & Pathologist

Remarks of ADG (Medical), CPMFs as Technical Approving Authority

Approved.

ADG (Medical), CPMFs