

No. IV-21011/45/2009-Prov-I
Government of India
Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House,
New Delhi, 29.7.2010

To, ^{11/04} ^{P.B.No-5/07} ^{11/02 5/08 5/07}
The DGs: Assam Rifles/BSF/CISF/CRPF/ITBP/NSG/SSB/BPR&D

Subject:- Specifications of various Medical/Hospital Equipments, Laboratory equipments and Laundry equipments -regarding


Sir,

The Specifications for the following Medical/Hospital Equipments, Laboratory equipments and Laundry equipments have been accepted and approved by the Competent Authority in MHA and the same are enclosed for information and Record:-

1. Condom Vending Machine- Annex-I
2. Flowcytometer- Annex-II
3. **Laboratory Equipments**
 - a) Coagulation Analyzer- Annex-III
 - b) Glycosylated Haemoglobin/Micro Albumin/CRP/D-Dimmer Analyzer-IV
 - c) Urine Analyzer – Annex-V
 - d) Semen Quality Analyzer- Annex-VI
 - e) Auto Pipette System- Annex-VII
 - f) Anaesthesia Work Station- Annex-VIII
- 4) **Laundry equipments**- Annex-IX
 - a) Washing Machine
 - b) Drier Washing Machine- 30 Kg (Moderate size)
 - c) Heavy Electric Press(Hand Operated)
 - d) Hydro Extractor
 - e) Table for pressing hospital linen
 - g) Sluice machine
 - h) Laundry Trolley
 - i) Bulk sterilizer electric for sterilizing linen and mattresses

2. Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Specifications.

Yours faithfully,



(S.B.Nanda)

Under Secretary to the Govt. of India

Encls: as above


Copy to:-DD(Procurement),MHA

Copy for information to:-PS to JS(PM),MHA


Technical Specifications of single column condom vending machine (coin operated)

1.	It should be mechanically operated
2.	It must have a storage capacity of 25 packets
3.	Option for wall mounting and table top installing
4.	It must be single columned condom vending machine
5.	The average dimensions of the machine should be 660mm * 330 mm * 255 mm (+ or - 15%)
6.	Body of the machine must be of atleast 2mm thick CRCA sheet chemically treated anti corrosion and rust free and powder coated
7.	Weight should be around 24 kgs (+ or - 15%)
8.	Must have sleek design and fine craftsmanship
9.	Must vend one pack of condoms on insertion of required coin as specified and subsequent operation of lever.
10.	There must be a provision for product/brand display on front/side panel of the machine
11.	Machine must accept only specified coin
12.	Fake coin and or coin of different denomination should be rejected by the machine and arrangements for taking back the coin through an outlet provided for that purpose must be there in the machine
13.	The machine must have provision for cabinet lock with padlocks
14.	There must be no electricity power requirement
15.	All internal parts must be sufficiently strong and rugged and duly surface treated
16.	There must be a provision for holding promotional material
17.	Must be designed to handle rough usage
18.	Coin box should be provided for coin collection and safe storage

Presiding officer


14/7/09
(Dr. N.K. Bhambri, IG (Med))

Member


14/7/09
(Dr. Amit Butola, CMO (S))


Member


14/7/09
(Dr. C.K.K Mao, CMO)

Member


14/7/09
(Dr. Sandhya Thapliyal, SMO)

Member


14/7/09
(Dr. Brijesh Kumar, CMO)

Appd.

Counter Signature


17/07/09

(Dr. R.S. Rathore)
ADG (Med), CPMFs