

No. IV-21011/9/2009-Prov-I
Government of India
Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House,
New Delhi, 18.5.2009

To


The DGs: Assam Rifles/BSF/CISF/CRPF/ITBP/NSG/SSB/BPR&D

Subject:- Technical Specifications/QRs of the Hospital Equipments

The Technical Specifications/QRs of the following Hospital Equipments as per Annexures, have been accepted by the Competent Authority in MHA:-

1	Digital Tabletop B Class Dental Autocalve
2	Digital X-Rays Machine for 50 beded Hospital
3	Advance Electro Therapy Unit, Combination Therapy Equipment with Electro Therapy with Ultrasound & Laser Therapy & Vacuum.
4	Whole Body Vibration Unit
5	High Frequency Shortwave Diathermy
6	Tens & Stimulator
7	Continuous Passive Motion(For upper Limb & Lower Limb)
8	Contrast bath(Hot & Cold)

2. Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Technical Specifications/QRs.


(R.S.Sharma)
Director (Prov)

Copy to:-

DD(Procurement),MHA

Copy for information to:-

PS to JS(PM),MHA

Copy also to:- The ADG(Med),
Central Paramilitary Force Medical Services,
Ministry of Home Affairs,
ITBP CASMP,
TIGRI,
P.O. Madangir,
New Delhi-110062.

ANNEXURE

QRs for Contrast bath(Hot & Cold)

The legendary treatment used for many years and widely distributed, bath of the upper and/ or lowers limbs make it possible to stimulate air circulation & ease the pain of rheumatoid type disorders	
MAIN FEATURES :	
1	Suitable for treatment of arms and legs.
2	Two Stainless steel heavy gauge tanks 15"x17"x19".
3	One tank fitted with special heating unit and other.
4	With heavy duty cooling unit. Each tank should well insulated and provided with thermostat for temperature control, pilot light and water outlet