

No. IV-21011/7/2010-Prov-I
Government of India
Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House,
New Delhi, 1.7.2010

To

1/10/10 *R.E.No - 16/09.*
01/01/10 *02/02* *16/09*
The DGs: Assam Rifles/BSF/CISF/CRPF/ITBP/NSG/SSB/BPR&D

Subject: - Specifications of various Medical/Hospital Equipments -regarding

Sir,

The Specifications for the following Medical/Hospital Equipments have been accepted and approved by the Competent Authority in MHA and the same are enclosed for information and Record:-

- 1) Glycosylated Hb Automatic Reader-Annex-A
- 2) Infant Open Care System-Annex-B
- 3) Refraction Unit-Annex-C
- 4) Auto Refractor-Annex-D
- 5) Embalming Machine-Annex-E
- 6) Oxygen Concentrator-Annex-F
- 7) Automated S. Electrolyte Reader-Annex-G
- 8) Dental Chair-Annex-H
- 9) Anaesthesia Work Station-Annex-I

2. Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Specifications.

Yours faithfully,

[Signature]
(S.B.Nanda)

Under Secretary to the Govt. of India

Encls: as above

Copy to:-

DD(Procurement),MHA

Copy for information to:-

PS to JS(PM),MHA

1
Annexure 'H'

TECHNICAL SPECIFICATIONS OF DENTAL CHAIR

GENERAL DESCRIPTION:

- Chair mount Dental Unit complete with all respects which should swivel left to Right and vice-versa for Left or Right handed Dentist.
- Should be full physiologic body contoured, completely motorized hydraulically driven Chair with absolute smooth non jerky, soft start and soft-stop movements.
- dental chair should be programmable with at least Four programmes of which two programmes for patient entry, one for rinse and one for patient exit, for two dentists A&B. One of the four programmes should be to set head low position in case of medical emergencies. Chair should have synchronised movement that when head rest goes down in supine or head low position the foot end of the chair board should raise. The chair should be programmable with a touch of a button.
- Chair should have possibilities of swivel left and right from base unit more than 30°
- Arm-rests should be multi-functional & easily adjustable, with at least one arm which rotates out of way to streamline patient entry/exit.
- Multi-purpose, adjustable head-rest for comfortable support of head, which should also be adjustable for pediatric & handicapped patients for exact adaption to the back of the head
- Touch pad control system should be multi-functional for Chair and Unit Master on/off. Halogen light on/off. with variable intensity, Bowl Flush, Tumbler fill, full chair function manual and programmable positioning.
- It should have functions for light cuce unit, intra oral camera, micro motor with endo function. All in built.
- In built Anti-retraction valves and flush valve system for infection control
- Multifunction, easy to use foot-switch. It should provide all chair movements, adjustable and programmable position
- Integrated 17 inches LCD/TFT monitor should be provided for RVG (brand should be mentioned)

Operating Light

- Should have an optically designed multi surface glass reflector having three intensity controls and with sensing device and shadow less.

Ascanth

J. S. S.

- 215
- Light Head should have three axial movements i.e Horizontal, Vertical and Axial diagonal adjustment at least 30 degree tilt on either side.
 - Should be quartz halogen Xenon bulb with spare bulb and holder located at Light Head area for quick maintenance.
 - Should have feathered- edge, balanced intensity light pattern.
 - Should have three position intensity switch with high medium and composite setting.

Suction and Syringes

- Should have high and low volume motorized Suction system with separating tank and automatic self-cleaning separating tank to allow continuous draining and equipped with fine particles collector and amalgam separators.
- Should have swivel Vitreous China Bowl with adjustable cup fill and Bowl rinse timers.
- Should have clean water bottle system
- Should have autoclavable Saliva ejectors, autoclavable high
- Volume evacuator and autoclavable six way syringe.

Ceramic Spittoon and Tumbler with water connection and auto start

Foot control:

Multifunctional, easy-to-use foot control. It should provide all chair movements, adjustable and programmable position

Dental working stool for dentist as well as for assistant:

- Should have cast-aluminium base with 5 -castors
- Should have adjustable lumbar support
- Integral hydraulic system for height adjustment

Modular Overhead delivery system with whipdown arms (Continental Delivery System) on Pneumatic brake arm

- Two terminals of Fiber-optic Head air rotor hand pieces
- One short head and one stand head air rotor hand pieces with spray and non-retraction valve.
- One micro motor (40,000 RPM) with fiber optic straight and contra hand pieces and one endo hand piece.
- Ultrasonic endo dentic hand piece: One complete set of Ultrasonic endo dentic tips and files.
- All hand-piece tubing should be of vinyl material.
- Unit block should be of pneumatic.
- No solenoid or kink valves.

Beaman

John

- All water line tubings should be made of ALPHASUN material. Tubing should have silver particle ionisation to protect water line

The following accessories will part of the equipment to be supplied.

Light curing unit LED-based with protective shield :

- It should be attached with chair unit
- Noise less & without fan
- Should be able to cure to a depth of 3-6 mm
- Should have microprocessor with three modes: fast, pulse and ramping.
- Should be without infra-red
- Light intensity should be able to generate optical power up to 2,200 mW/cm²
- Wavelength range 420-480nm

Piezo electric scaler:

- It should also be provided in overhead delivery module.
- It should have automatic protection control for maximum safety
- Ultra-sound frequency ranging from at least 27,000-33,000 Hz
- Should be provided with frequency tuning.
- Should have an intelligent control system of amplitude according to the resistance encountered by the tip and the setting mode selected.

Instrument table:

- Stainless Steel removable instrument trays (quantity - 2 nos.)
- Must have swing-arm with wide swivel
- Size not less than 20' x 8'
- With height adjustment and locking facility

Oil free compressor:

- It should be compact and silent with internal anti-corrosive reservoir
- It should be One cylinder made up of cast iron, oil free with dry air system
- It should have tank capacity of at least 20 litres
- It should have auto cut-off switch
- Pressure range: 5.5 - 8.0 bar
- Should have suction capacity of 100 to 120 litres per minute at least.

X-ray viewer: for viewing ICDA x-ray must be in-built with the unit

RVG unit including intra-oral X-ray machine, computer, USB cam (high resolution digital Intra-oral camera):

Masamhan

Subin

4

Intra-oral X-ray machine, micro-processor controlled, compatible to
RVC

- Tube current: 70 kv
- Focal spot: 0.8 mm.
- Focus to skin distance: standard 20 cm
- Duty cycle: At least 1:32, i.e. waiting time between two exposures for cooling of X-ray tube (10-20 sec at least)
- Timer settings: Child/adult selection, upper & lower jaw selection, also Bitewing & occlusion function
- Exposure times: range from 0.02 to 3.2 sec at least
- Low-Dose setting for digital receptor
- Facility of remote operation should be provided.
- Automatic correction facility for mains fluctuations
- Hand switch timer with at least 3 mtr coiled cord
- Hygienic touch- buttons
- Radiation safety apron should be provided
- Double pantographic arm with vertical & horizontal smooth movements, floor mount, mobile on castors,
- The system should comply with national quality certification /international standards for radiation safety including CE/BARC/AERB certified

Digital intra-oral camera should have following specification -

- Should be portable, light weight, convenient and user friendly
- Should be card free & should have USB cable connectivity
- Should have focus free optics & fog free

R V G SENSOR & SOFTWARE -

- Theoretical resolution: more than 25 lines per mm.
- Actual true resolution: more than 20 lines per mm
- Gray shades that can be resolved: 4000 minimum
- Number of pixels in sensor: 1.8 pixel minimum
- The product should be totally mobile and connected to PC through USB
- Sensor should have smooth surfaces with rounded corners & edges for patient comfort
- Should have image sharpness filters
- It should have interface features like direct integration with other devices like intra-oral camera, panoramic systems, scanners, etc.
- Conversion/upgradation of the software should be available
- Length of sensor wire should be minimum 2 metres
- Sensor should be provided with software facility to enhance zoom, colorize, invert & rotate the image.
- Sensor life >400,000 exposures
- Should have auto take feature - no need to reset between exposure
- Should be DICOM compatible, user-friendly and powerful program that acquires processes and archives images.

Baamle

Jub. 11/11

- Should have caries detection software.
- Should be supplied with sensor holder.
- COMPUTER with latest configuration (256 GB HDD, 2 GB RAM, Intel core2 duo processor >2.5 GHz, with windows 7 OS).
- 17" TFT screens (one mounted on chair and other on desk) & colour printer with appropriate cabinet for computer and printer (Make/model of the computer/ peripherals should be mentioned)

Endodontic micro motor with built in apex locator:

- Electronically controlled, endodontic micro-motor with built in apex locator
- Should start automatically once file enters the canal & stop automatically once it reaches preset position near apex
- Should stop & reverse if the file clogs in canal
- Should negotiate curved canals
- Should have high, low & manual speed modes with auto-torque detection
- Autoclavable hand piece, mini. contra-head for easy access & viewing

Glass bead sterilizer (ISI certified) for disinfecting endodontic files with appropriate ratings

Digital -Electronically Controlled steamfast AUTOCLAVE Table top type

- Micro-processor controlled electronics
- Patented steam generation system
- Capacity 16-20 litres at least
- Customized vacuum drying
- Wide range of sterilization programmes
- Also should have an option of emergency sterilization program of at least 10 minutes lesser than this is desirable.
- Integrated water filling & draining tanks
- Electronic-controlled locking system for additional safety

Stabilizer:

- High quality, ISI certified, with appropriate ratings for above operatory

HAND PIECE STERILISER

- UV hand piece steriliser system with automatic 35 seconds maintenance programme for cleaning and lubrication of all types of hand pieces

Dental Ultrasonic cleaning Unit Bench Top

- Bench Top cleaner with electronic timer water resistant construction, stainless steel tank, with 2 to 3 litres capacity

Instrument cabinets -

- one Sterile UV light storage cabinet for autoclaved instruments

Prasanna

JS

- one dental instrument stackable cabinet having wash basin, top made of artificial marble/granite, durable, easy to clean with at least 3-4 separate drawer compartments, with inner power supply socket to use small equipments.
- One mobile stackable cabinet should also be provided.

System configuration accessories, spares and consumables:


- All consumables for installation/standardization of the system should be given free of cost
- Complete installation of the system including water input and drainage has to be installed


Standards and safety:


The accessories of the dental unit should carry appropriate certifications like FDA/CE/BIS/ISO/ IEC/ISI, whichever applicable. Make & model of all accessories/attachments should be specified


Documentation:


- Certificate of calibration and inspection
- User /technical/maintenance manuals to be supplied
- List of accessories with their number and costing
- Log Book with instructions for maintenance purpose.



DR.N.k.Bhanbri
IG (Medica)
PO


Dr. Brijesh Kumar
2IC (Medical) SSB
MEMBER-I



Dr. Amit Butola
CMO BSF
MEMBER-II


Dr. Sandhya Thapliyal
Specialist MO/DC, ITBP
MEMBER-III


Dr. Jitendra Kumar Sahu
MO CRPF
MEMBER-IV


Dr. Zabceulia
AC DENTIST, ITBP
CO-OPTED MEMBER

TECH. APPROVED


DR. R.S Rathore
ADG (MEDICAL)