

No. IV-21011/9/2009-Prov-I
Government of India
Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House,
New Delhi, 18.5.2009


To
The DGs: Assam Rifles/BSF/CISF/CRPF/ITBP/NSG/SSB/BPR&D

Subject:- Technical Specifications/QRs of the Hospital Equipments

The Technical Specifications/QRs of the following Hospital Equipments as per Annexures, have been accepted by the Competent Authority in MHA:-

1	Digital Tabletop B Class Dental Autoclave
2	Digital X-Rays Machine for 50 beded Hospital
3	Advance Electro Therapy Unit, Combination Therapy Equipment with Electro Therapy with Ultrasound & Laser Therapy & Vacuum.
4	Whole Body Vibration Unit
5	High Frequency Shortwave Diathermy
6	Tens & Stimulator
7	Continuous Passive Motion(For upper Limb & Lower Limb)
8	Contrast bath(Hot & Cold)

2. Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Technical Specifications/QRs.


18/05/09
(R.S.Sharma)
Director (Prov)

Copy to:-

DD(Procurement),MHA

Copy for information to:-

PS to JS(PM),MHA

Copy also to:- The ADG(Med),
Central Paramilitary Force Medical Services,
Ministry of Home Affairs,
ITBP CASMP,
TIGRI,
P.O. Madangir,
New Delhi-110062.

ANNEXURE

QRs FOR DIGITAL TABLE TOP "B" CLASS DENTAL AUTOCLAVE

1	Dimensions: 594 X522x410 (mm)
2	Weight: 49 Kgs
3	Option of Dry vacuum " Autoclaving"
4	Auto refill water system
5	Chamber size 230x410 (mm)
6	Water Chamber capacity 4200 cc
7	Voltage 230V
8	Power consumption 1900 watts/08 Amp
9	Provision of pressure safety valve
10	Electric over power protection 20 Amp fuse
11	Auto dry time setting at 0,15,30,35 & 40 minutes
12	Provision of bovine testing for sterilization
13	Provision of Chamber water level sensor
14	Provision of pressure indicator and display
15	Provision of pressure sensor
16	Provision of temperature sensor
17	Provision of temp. display
18	Loading/packing 3 X tray, wrapping/unwrapping option, dry/wet option
19	Distilled water Plant: Distilled water generating plant to included as a standard accessory/pout alongwith the autoclave.