## No. IV-21011/7/2010-Prov-I Government of India Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House, New Delhi, 1.7.2010

To	. L.	0.8.10- 16/09.		
111	سب سلاما		G	11609
The DGs:Assar	n Rifles/BS	F/CISF/CRPF/ITE	BP/NSG/SŚB/P	PR&D

Subject: - Specifications of various Medical/Hospital Equipments -regarding

Sir,

The Specifications for the following Medical/Hospital Equipments have been accepted and approved by the Competent Authority in MHA and the same are enclosed for information and Record:-

- Glycosylated Hb Automatic Reader-Annex-A
   Infant Open Care System-Annex-B
   Refraction Unit-Annex-C

- Netraction Unit-Annex-C
   Auto Refractor-Annex-D
   Embalming Machine-Annex-E
   Oxygen Concentrator-Annex-F
   Automated S. Electrolyte Reader-Annex-G
   Dental Chair-Annex-H
   Anaesthesia Work Station-Annex-I

Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Specifications.

Under Secretary to the Govt. of India

Encls: as above

Copy to:-

DD(Procurement),MHA

Copy for information to:-

PS to JS(PM),MHA

Annexus F

## Appropris II

## Technical specifications for Embalming Machine

- 1. Fluid delivery rate should be 10 ltrs /hr
- Inner tank to store embalming fluid with capacity 5-7 ftrs. should be of stain less steel
- 3. Pump: pump should be of ELECTROMAGNETIC DOSING PUMP with capacity 0 5 ftrs per Hour and pressure 3kg/cm2
- 4. The equipment should be mounted on castors for easy movement and the grip should be provided for lifting
- 5. I.V stand fixed for mounting cannula tubing and mains cable.
- 6. Indicator for mains on & in use should be present
- 7. The outer body should be of complete stainless steel.

Presiding officer	Bearly.
Member BSF	(Dr. N.K. Bhambri, IG (Med))
Member CRPF	(Dr. Rohit Kumar., CMO, BSF)
Member ITBP	(Dr. dyc.KM/63, OMO, CRPF)
Member SSB	(Dr. D.K Verma, SMO)  (Dr. B.K. Nigam, CMO (SG))

Dr. B. Singh, MS (Gon. Surgery), CMO (SG), ITBP)

Technically approved/not approved

ADG (Med), CPFs