

No. IV-21011/45/2009-Prov-I
Government of India
Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House,
New Delhi, 29.7.2010

To, ^{11/04} ^{P.B.No-5/07} ^{11/02 5/08 5/07}
The DGs: Assam Rifles/BSF/CISF/CRPF/ITBP/NSG/SSB/BPR&D

Subject:- Specifications of various Medical/Hospital Equipments, Laboratory equipments and Laundry equipments -regarding


Sir,

The Specifications for the following Medical/Hospital Equipments, Laboratory equipments and Laundry equipments have been accepted and approved by the Competent Authority in MHA and the same are enclosed for information and Record:-

1. Condom Vending Machine- Annex-I
2. Flowcytometer- Annex-II
3. **Laboratory Equipments**
 - a) Coagulation Analyzer- Annex-III
 - b) Glycosylated Haemoglobin/Micro Albumin/CRP/D-Dimmer Analyzer-IV
 - c) Urine Analyzer – Annex-V
 - d) Semen Quality Analyzer- Annex-VI
 - e) Auto Pipette System- Annex-VII
 - f) Anaesthesia Work Station- Annex-VIII
- 4) **Laundry equipments-** Annex-IX
 - a) Washing Machine
 - b) Drier Washing Machine- 30 Kg (Moderate size)
 - c) Heavy Electric Press(Hand Operated)
 - d) Hydro Extractor
 - e) Table for pressing hospital linen
 - g) Stoice machine
 - h) Laundry Trolley
 - i) Bulk sterilizer electric for sterilizing linen and mattresses

2. Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Specifications.

Yours faithfully,



(S.B.Nanda)

Under Secretary to the Govt. of India

Encls: as above

Copy to:-DD(Procurement),MHA

Copy for information to:-PS to JS(PM),MHA

**TECHNICAL SPECIFICATIONS OF GLYCOSYLATED HAEMOGLOBIN/
MICROALBUMIN/CRP/D-DIMMER ANALYZER**

1. System should be portable & easy to carry.
2. It should have the facilities to measure Four different Tests like HbA1c, U-Albumin, C-Reactive Protein & D-Dimer Quantitative Estimation
3. It should have the facility to conduct test in a very short time
4. It should be compact and light weight
5. It should have the facility of display on LCD
6. It should be provided with Rechargeable battery with battery Charger
7. Guarantee/Warranty/AMC must be ensured.

Presiding officer	<u>M. N. K. Bhambri</u> Dr. N.K. Bhambri IG(Med)/Medical Supdt. of Referral Hospital
Member from BSF	<u>Rohit Kumar</u> Dr. Rohit Kumar, CMO
Member from CRPF	<u>Bipul Kumar</u> Dr. Bipul Kumar, MO
Member from ITBP	<u>Sandhya Thapliyal</u> Dr. Sandhya Thapliyal, SMO
Member from SSB	<u>Brijesh Kumar</u> Dr. Brijesh Kumar, CMO (SG)
Co-op Member	<u>B.S. Mishra</u> Dr. B.S. Mishra IG (Medical) & Pathologist

Remarks of ADG (Medical), CPMFs as Technical Approving Authority

Approved.

ADG (Medical), CPMFs

25/09/09