

No. IV-21011/7/2010-Prov-I
Government of India
Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House,
New Delhi, 1.7.2010

To
The DGs: Assam Rifles/BSE/CISF/CRPF/ITBP/NSG/SSB/BPR&D

11/06 *RE.No-16109* *16/09*

Subject: - Specifications of various Medical/Hospital Equipments -regarding

Sir,

The Specifications for the following Medical/Hospital Equipments have been accepted and approved by the Competent Authority in MHA and the same are enclosed for information and Record:-

- 1) Glycosylated Hb Automatic Reader-Annex-A
- 2) Infant Open Care System-Annex-B
- 3) Refraction Unit-Annex-C
- 4) Auto Refractor-Annex-D
- 5) Embalming Machine-Annex-E
- 6) Oxygen Concentrator-Annex-F
- 7) Automated S. Electrolyte Reader-Annex-G
- 8) Dental Chair-Annex-H
- 9) Anaesthesia Work Station-Annex-I

2. Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Specifications.

Yours faithfully,

(S.B.Nanda)

Under Secretary to the Govt. of India

Encls: as above

Copy to:-

DD(Procurement),MHA

Copy for information to:-

PS to JS(PM),MHA

SPECIFICATION OF REFRACTION UNIT

Annexure C

Examination Chair Unit, Ophthalmic Refraction Chair Unit

- Basic Standard integrated unit contains:
- One fully upholstered Elegant Ophthalmic Chair
 - With full motorized recline facilities
 - With full motorized up & down movements for 300mm One Stand and Console
 - With illuminating soft light for examination
 - With controls for Ophthalmic Chair
 - For "recline" and "up & down"
 - For "up & down" movement foot pedal is also provided as an additional option.
 - With motorized vision drum
 - With Provision for fitting following optional instruments
 - * Slit lamp
 - * Binocular Indirect Ophthalmoscope
 - * Keratometer OR Auto-Refractometer
 - * Direct Ophthalmoscope and Steak Retinoscope
 - * Chart Projector
 - * Trial Lens Set
 - * Sight Test (Phoropter).

TECHNICAL SPECIFICATIONS

Base Dimension	Floor space required
Height	7.6'
Length	After reclining 9.6'
Width	4'
Input voltage & power	110/230V
Output	3V, 6V, 12V
Weight	80Kg (approx)

[Signature]
P.O.

Dr H.S. Dharam Shetty
SMD
ITRP

[Signature]
MI

Dr DK Verma
DC / Eye Specialist
ITRP

[Signature]
MD

Dr Sundhya Th
DC / Ophthalmologist
ITRP

[Signature]
05/05/89
Dr. H. S. Dharam Shetty
Aggr. D. O. (Medical), C.P.S.
ITRP, Convent, Tagh M, Delhi-62