

No. IV-21011/7/2010-Prov-I
Government of India
Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House,
New Delhi, 1.7.2010

To

11/06 *P.B.No - 16/09*
05/07 *16/09*
The DGs: Assam Rifles/BSF/CISF/CRPF/ITBP/NSG/SSB/BPR&D

Subject: - Specifications of various Medical/Hospital Equipments -regarding

Sir,

The Specifications for the following Medical/Hospital Equipments have been accepted and approved by the Competent Authority in MHA and the same are enclosed for information and Record:-

- 1) Glycosylated Hb Automatic Reader-Annex-A
- 2) Infant Open Care System-Annex-B
- 3) Refraction Unit-Annex-C
- 4) Auto Refractor-Annex-D
- 5) Embalming Machine-Annex-E
- 6) Oxygen Concentrator-Annex-F
- 7) Automated S. Electrolyte Reader-Annex-G
- 8) Dental Chair-Annex-H
- 9) Anaesthesia Work Station-Annex-I

2. Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Specifications.

Yours faithfully,

(S.B.Nanda)

Under Secretary to the Govt. of India

Encls: as above

Copy to:-

DD(Procurement),MHA

Copy for information to:-

PS to JS(PM),MHA

SPECIFICATION OF REFRACTION UNIT

Annexure - C

Examination Chair Unit; Ophthalmic Refraction Chair Unit

Basic Standard integrated unit contains:

One fully upholstered Elegant Ophthalmic Chair

-- With full motorized recline facilities

-- With full motorized up & down movements for 300mm

One Stand and Console

-- With illuminating soft light for examination

-- With controls for Ophthalmic Chair

For "recline" and "up & down"

For "up & down" movement foot pedal is also provided as an additional option.

With motorized vision drum

-- With Provision for fitting following optional instruments

* Slit lamp

* Binocular Indirect Ophthalmoscope

* Keratometer OR Auto-Refractometer

* Direct Ophthalmoscope and Steak Retinoscope

* Chart Projector

* Trial Lens Set

* Sight Test (Phoropter).

TECHNICAL SPECIFICATIONS

Base Dimension

Floor space required

Height

: 7.6'

Length

: After reclining 9.6'

Width

: 4'

Input voltage & power

: 110/230V

Output

: 3V, 6V, 12V

Weight

: 80Kg (approx)

P.O.

Dr. DK Verma
MI

Sandhya
MB

Dr. H.S. Dharam Singh
SMD
ITBP

Dr. DK Verma
DC / Eye Specialist
ITBP

Dr. Sandhya Th
MB / Gynaecologist
ITBP

Approved
55/05/89
भारत वाणिज्यिक (चिकित्सा) कंपनी
Adm. D. O. (Medical), CPFS
एन.डी.ए. कैंपस, रिजर्व गेट फेज-52
Hbr Camps, Tugh N, Delhi-62