No. IV-21011/9/2009-Prov-I Government of India Ministry of Home Affairs

26, Man Singh Road, Jaisalmer House, New Delhi, 18.5.2009

To

The DGs: Assam Rifles/BSF/CISF/CRPF/ITBP/NSG/SSB/BPR&D

Subject:- Technical Specifications/QRs of the Hospital Equipments

The Technical Specifications/QRs of the following Hospital Equipments as per Annexures, have been accepted by the Competent Authority in MHA:-

_1	Digital Tabletop B Class Dental Autocaive
2	Digital X-Rays Machine for 50 heded Hospital
3	Advance Electro Therapy Unit, Combination Therapy Equipment with Electro Therapy with Ultrasound & Laser Therapy & Vacuum.
4	Whole Body Vibration Unit
5	High Frequency Shortwave Diathermy
6	Tens & Stimulator
7	Continuous Passive Motion(For upper Limb & Lower Limb)
8	Contrast bath(Hot & Cold)

Henceforth, all the CPMFs should procure the above items required by them strictly as per the laid down Technical Specifications/QRs.

(R.S.Sharma)

Director (Prov)

Copy to:-

DD(Procurement),MHA

Copy for information to:-

PS to JS(PM), MHA

Copy also to:- The ADG(Med),

Central Paramilitary Force Medical Services,

Ministry of Home Affairs, ITBP CASMP,

TIGRI,

P.O. Madangir, New Delhi-110062.

<u>ANNEXURE</u>

ORs for Whole Body Vibration Unit

i =	Physiotherapy
1	- Arthrosis
	-Circulatory trouble
	-Varicose veins
	-Requirculation
ii	Osteoporosis
iii	The Linit creates special months in 1.7
ív	The Unit creates special mechanical vibrations and transmits them to the body More than 200 protocols with images
v	Touch screen display
vi	Possibility to annual to
vii	Possibility to program and personalize the parameters Frequency 10-80 Hz
viii	Villantia III
íx	Vibration Amplitude; Low~ high
	No. of motors: Two
j eenn	cal specifications:-
ii .	Main supply: - 240,220,120,120CV AC
	Out frequency:- 27,12 MHz
iii	Output Power: 0-375 W(Continuous) 150W in peak Power in Pulsed Mode
iv	Output Modes:- Continuous Pulsed 3/3:2/3:1/3
v	Pulse widths:-20/40/65/100/200/400 microsecond
vi	Pulse repetition rates: 50-800 pulses per second
vii	Automatic tuning
víii 🗍	Electrodes: - Rubber pad, Condenser electrodes
ix _	Timer: 00-29 minutes
x T	Clear display and touch - sensitive controls
ki 💮	A range of electrodes available for all types of treatments