

Government of India
Ministry of Home Affairs
PM Division (Procurement Wing)
Jaisalmer House, 26 Man Singh Road
New Delhi-110011.
Tel: 91-011-23387306.
Email: inder.jeet73@gov.in

Dated, the 13 February, 2020.

CORRIGENDUM NO.1
TO GLOBAL TENDER ENQUIRY (GTE)
No. D/21013/3291/19.11.2019/PW

REF: GLOBAL TENDER ENQUIRY NO.D/21013/3291/19.11.2019/PW DATED 23.01.2020 FLOATED ON CPP PORTAL AND SCHEDULED TO BE OPENED ON 21.02.2020 FOR PROCUREMENT OF PORTABLE RAMAN SPECTROMETER Qty. 01 UNIT FOR PHYSICS DIVISION OF LNJNI NICFS, ROHINI NEW DELHI.

A Pre Bid Meeting under the chairmanship of DIG (Admin), LNJNI NICFS was held on 31.01.2020 at 11.00 AM at LNJNI NICFS, Rohini Sector-3, New Delhi to address the points raised by the prospective bidders in connection with Global Tender Enquiry (GTE) issued by MHA vide NO.D/21013/3291/19.11.2019/PW dated 23.01.2020 floated on CPP Portal for procurement of "Portable Raman Spectrometer System" Qty. 01 unit. After discussion in the pre bid meeting, following amendments are hereby issued against the subject GTE. These changes may be effected wherever it is defined in the text of the tender documents and it may be treated as amended accordingly:-


(Technical Parameters of GTE)

S.No.	Ref. of Para in GTE	Clarification Sought/ suggestion submitted by Rep. of firms	Changes agreed/ Agreed by amendment in GTE	Not MHA/
1.	Appendix-2 S.No. 1 at Page No. 40 of GTE Existing Clause:- 1. For 532nm excitation, (50mW max.)	M/s Lab India Instruments Pvt. Ltd. requested that the value should be > 35m W at Existing Port	For Existing Clause To be read as:- "For 532nm excitation, (between 35mW to 50mW).	
2.	Appendix-2 S.No. 6 at Page No. 40 of GTE	M/s Lab India Instruments Pvt. Ltd. requested that	For Existing Clause To be read as:-	

13/02/20

	<p>Existing Clause:- 6. System should have resolution of 4.5 cm-1 and detector should be CCD array with 2058 Pixel at least.</p>	<p>the value should be 2048 pixel detector.</p>	<p>"System should have resolution of 4.5 cm-1 and detector should be CCD array with 2048 Pixel at least."</p>
3.	<p>Appendix-2 S.No. 9 at Page No. 40 of GTE</p> <p>Existing Clause:- 9. Direct 21 CFR Part 11 compliancy software (quote optionally namely (i) Data Library for Narcotics & Drugs & (i) Data Library for Explosives.</p>	<p>M/s Lab India Instruments Pvt. Ltd. requested that the libraries are available individually and not as a bunch. Therefore each optional library may be quoted for more clarity.</p>	<p>Agreed by the Committee</p> <p>For Existing Clause</p> <p>To be read as:- "Direct 21 CFR Part 11 compliancy software (quote optionally each namely (i) Data Library for Narcotics & Drugs & (ii) Data Library for Explosives)."</p>
4.	<p>Appendix-2 S.No. 10 at Page No. 40 of GTE</p> <p>Existing Clause:- 10. Raman Data/software interface to quantitative analysis.</p>	<p>M/s Lab India Instruments Pvt. Ltd. requested that The data base available is able to do relative quantitative analysis therefore relative word may be added for more clarity.</p>	<p>Agreed by the Committee</p> <p>For Existing Clause</p> <p>To be read as:- "Raman Data/software interface to relative quantitative analysis."</p>

2. All other terms and conditions will remain unchanged. The Bidders may submit their bid accordingly.


(Inderjeet Chawla)
2IC (Proc) MHA

For and on Behalf of President of India.

Copy to:

1.	The Director, LNJN NICFS, Outer Ring Road, Rohini Sector-3, New Delhi-110085.	For information with reference to this wing GTE of even no. dated 23.01.2020
2.	Ministry of External Affairs, UOI, New Delhi.	With request to forward copy of corrigendum notice to the Ambassador/ High Commissioner of Embassies of those countries who are likely suppliers of subject store.
✓ 3.	The Section Officer (IT), IT Cell Website Room No. 10, MHA, North Block, New Delhi-110001.	With request to get it incorporate on the official website of the Ministry of Home Affairs.

Chand
13/2/2020