

**GOVERNMENT OF INDIA  
MINISTRY OF HOME AFFAIRS  
(PROCUREMENT WING)  
JAISALMER HOUSE, 26 MAN SINGH ROAD,  
NEW DELHI-110 011**

**DATED 26.12.2018**

**CORRIGENDUM NO. 01  
TO GLOBAL TENDER ENQUIRY (GTE)  
No. D/21013/3280/09.11.2018/P-1 DATED 05.12.2018**


**Ref: GLOBAL TENDER ENQUIRY (GTE) NO. D/21013/3280/09.11.2018/P-1 DATED 05.12.2018 FLOATED ON CPP PORTAL AND SCHEDULE TO BE OPENED ON 08.01.2019 (TECH BID) FOR SUPPLY, INSTALLATION AND COMMISSIONING OF THERMO CYCLERS/PCR QTY. 18 UNITS FOR ESTABLISHMENT OF STATE OF ART FORENSIC LABORATORY AT CFSL, CHANDIGARH UNDER DFSS,**

A pre-bid meeting under the Chairmanship of Director, DFSS, New Delhi was held on 13.12.2018 at 1130 hrs to address the points raised by prospective bidders in connection with Global Tender Enquiry (GTE) issued by MHA on 05.12.2018 floated on CPP portal for procurement of Thermo Cyclers/PCR - Qty. 18 Nos. for CFSL, Chandigarh of DFSS. After discussion in the pre-bid meeting, after approval of competent authority, following amendments are hereby issued against the subject GTE. These changes may be effected wherever it is defined in the text of the tender documents and it may be treated as amended accordingly:-

**(Minutes of Pre-Bid Meeting)**

| SNo. | Ref. of Para in GTE  | Clarification Sought by Rep. of firms                           | Changes agreed/Not agreed by MHA/Amendment in GTE   |
|------|--|---|---|
| 1    | <b>Appendix-II (General Specification) at page No.35 of GTE.</b><br><br><b>Existing Clause :-</b><br><br><b>Sl.No.2</b> Temperature range 4°C to <b>99.9°C</b> with accuracy $\pm 0.25^\circ$ C or better. | <b>M/s Eppendorf India Ltd.</b> requested to clarify the clause | Amendment:<br>For:<br>Existing entries<br><b>Read :</b><br><br><b>Sl.No.2</b> Temperature range 4°C to <b>99°C</b> or more with accuracy $\pm 0.25^\circ$ C or better |
|      | <b>Sl.No.7</b> Touch screen colour display of temperature profiles & set up parameters.  |   | <b>Sl.No.7</b> Touch screen colour display <b>or Button operated display</b> of temperature profiles & set up parameters.   |
|      | <b>Sl.No.9</b> Validated for forensic DNA analysis as per SWGDAM guidelines.   |   | <b>Sl.No. 9</b> System should be capable of amplifying DNA using available kits for Forensic DNA analysis in the market.  |

2. All other terms and conditions will remain unchanged. The bidders may submit their bids accordingly.

  
(Rajendra Singh) DC(Proc)  
**For and on behalf of the President of India**

copy to: -

|   |  |   |
|---|--|---|
| 1 | <b>Indenter</b> - Director, CFSL, DFSS,<br>Block No 9, 8 <sup>th</sup> Floor, CGO Complex,<br>Lodi Road, New Delhi 110003. | For information with reference to this office GTE of<br>even No 05.12.2018.   |
| 2 | Ministry of External Affairs, UOI<br>New Delhi.  | With requested to forward copy of corrigendum<br>notice to the Ambassador / High commissioner of<br>Embassies of those countries who are likely<br>suppliers of subject item. |
| 3 | The Section Officer (IT), IT Cell<br>Website room no. 10, MHA, North<br>Block, New Delhi-110001.                           | With requested to get it incorporated in the official<br>website of the Ministry of Home Affairs.   |
| 4 | Name & address of Consignees   | Director, Central Forensic Laboratory, Plot No. 2,<br>Dakshin Marg, Sector 36-A, Chandigarh-160036.   |