

GOVERNMENT OF INDIA
MINISTRY OF HOME AFFAIRS
(PROCUREMENT WING)
JAISALMER HOUSE, 26 MAN SINGH ROAD,
NEW DELHI-110 011

Dated 27.12.2018

CORRIGENDUM NO. 01
TO GLOBAL TENDER ENQUIRY (GTE)
No. D/21013/3279/22.11.2018/P-1 DATED 06.12.2018

REF: GLOBAL TENDER ENQUIRY (GTE) NO. D/21013/3279/22.11.2018/P-1 DATED 06.12.2018 FLOATED ON CPP PORTAL AND SCHEDULE TO BE OPENED ON 09.01.2019 (TECH BID) FOR REGARDING OF PROCUREMENT OF 07 SYSTEM OF HIGH RESOLUTION SPECTRAL COMPARATOR FOR SAUS OF CFSL AT NEW DELHI, MUMBAI, & CHENNAI, & DFSS NEW DELHI AT CHANDIGARH, SHIMLA & HYDERABAD AND LNJN, NICFS, ROHINI.

A pre-bid meeting under the Chairmanship of Director, DFSS, New Delhi was held on 14.12.2018 at 1100 AM to address the points raised by prospective bidders in connection with Global Tender Enquiry (GTE) issued by MHA on 06.12.2018 floated on CPP portal for procurement of 07 System of High Resolution Spectral Comparator for SAUS of CFSL at New Delhi, Mumbai & Chennai, & DFSS New Delhi at Chandigarh, Simla & Hyderabad and LNJN, NICFS, Rohini, Delhi. After discussion in the pre-bid meeting, after approval of competent authority, following amendments are hereby issued against the subject GTE. These changes may be effected wherever it is defined in the text of the tender documents and it may be treated as amended accordingly:-

(Minutes of Pre-Bid Meeting)

SNo.	Ref. of Para in GTE	Clarification Sought by Rep. of firms	Changes agreed/Not agreed by MHA/Amendment in GTE
1	Annexure-IV Sl.No.3 at page No.31 of GTE. Existing Clause:- Training after installation and commissioning – The seller has to provide free operational training for seven working days to at least two Scientists at the time of installation as required by the Scientists at consignee locations. This will also include two sets of	M/s Perceptive Biotech Pvt Ltd. Can the training be conducted at any one of the consignees' locations for all the consignees in one go? This would help to complete the training of seven working days within stipulated	For Existing clause To be read as :- Training after installation and commissioning – The seller has to provide free operational training for three working days to at least two Scientists after

M

	<p>operational manual. Two days additional training for two scientists has to be provided in case of up gradation of Hardware/software, if any.</p>	<p>time? Since the final disbursement is linked to completion of training, imparting the training at each locations would consume close to or more than sixty days (including travel of the trainer) and it would have direct impact on the final disbursement accordingly.</p> <p>Firm requested to clarify the clause.</p>	<p>installation as required by the Scientists at consignee locations. This will also include two sets of operational manual. Two days additional training for two scientists has to be provided in case of up-gradation of Hardware/software or to address any scientific and technical issues related with operation of equipment, if any.</p>
2	<p>Annexure-IV Sl.No.5 at page No.32 of GTE.</p> <p>Existing Clause:-</p> <p>Inspection Methodology Joint Receipt Inspection of the stores shall be carried out by an inspecting authority / board so detailed by the respective Consignees i.e.</p> <p>(i) Director CFSL, CBI Block No-4 CGO Complex, Lodi Road, New Delhi 110003.</p> <p>(ii) Director DFSS, Block 9, 8th floor CGO Complex, Lodhi Road , New Delhi 110003</p> <p>(iii) Director, LNJN National Institute of Criminology & Forensic Science, Sector-3, Rohini, Delhi .</p> <p>The Indenters will also indicate the tentative date, time and place for such inspection.</p>	<p>M/s Perceptive Biotech Pvt Ltd.</p> <p>Whether all seven units would be inspected at one location before despatching to respective locations or would the inspection be carried at respective locations? The consignments are always packed with intercontinental air cargo worthiness. Once up packed, repacking them for onward delivery to distant locations is not recommended. Hence this query</p>	<p>Tender Condition Persists. As all are separate indenters</p>

Re

<p>3.</p>	<p>Appendix-2 Sl.No.1 at page No.38 of GTE.</p> <p>General Specifications Existing Clause:-</p> <p>High-resolution Hyper-spectral comparator System with High Resolution CCD/ digital Fire Wire Color Camera with suitable resolution in Megapixels to have high quality image to avoid blurry and pixelated image, Camera Spectral Response range approx. 400 or less to 1000nm or more.</p>	<p>M/s Perceptive Biotech Pvt Ltd. & M/s Vajra Bio Metric Pvt. Ltd.,</p> <p>Can the specification of System with High Resolution CCD/digital Fire Wire Camera be generalized to latest digital camera, instead of naming a particular camera? Since different manufacturers use different cameras suitable to achieve the results. Naming a particular type of Camera may affect the compliance for the bidders whose OEM use a Camera by different name. A generic term viz. latest Digital Camera would provide level playing field to all prospective bidders.</p>	<p>For Existing clause</p> <p>To be read as :-</p> <p>High-resolution Hyper-spectral comparator System with High Resolution CCD/digital Color Camera Minimum 5 Megapixels capable of successfully and suitably complying all the technical functions required to be rendered by the equipment as mentioned to have high quality image display in full HD or better to avoid blurry and pixelated image, Camera Spectral Response range approx.. 400 or less to 1000nm or more.</p>
<p>4.</p>	<p>Appendix-10 BOQ</p> <p>Existing Clause:- BOQ_435763.xls</p> <p>Price bid in Excel sheet as well as PDF format.</p>	<p>M/s Vajra Bio Metric Pvt. Ltd.,</p> <p>As you have already provided the excel sheet and request to please clarify on PDF format.</p>	<p>It is clarified that Price should be quoted in Excel sheet only .</p> <p>However , it may be read as Excel Sheet only wherever it is mentioned in GTE.</p>


Re

5.	<p>Annexure-II Sl.No. 13 at page No.19 of GTE</p> <p>Existing Clause:- SYSTEM OF BIDDING:</p> <p>SINGLE STAGE IN TWO BID SYSTEM</p> <p>Bidders will submit their techno-commercial bids and price bids online on the CPP Portal . The website also has user manuals with detailed guideline on enrolment and participation in the online bidding process. No conditional bid shall be allowed/accepted. All pages of tender should be numbered and properly indexed. Bidders will have to upload scanned copies of various documents required for eligibility and all other documents as specified in GTE, techno-commercial bid in cover-I, and price bid in cover-II . PRICE BID for the 'HIGH RESOLUTION SPECTRAL COMPARATOR SYSTEM' and CAMC should be scanned and uploaded online only.</p>	<p>M/s Vajra Bio Metric Pvt. Ltd., Price Bid for the High Resolution Spectral Comparator System and CAMC should be scanned and uploaded online only.</p> <p>Please clarify.</p>	<p>Tender Condition persists.</p> <p>Price bid to be up-loaded as per BOQ at Appendix-10.</p>
6	<p>Annexure-II Sl.No. 13.2 at page No.19 of GTE</p> <p>Existing Clause:-</p> <p>COMMERCIAL/PRICE BID (second cover):</p> <p>Price should be quoted in second over online only in prescribed format as per Appendix-10(Excel sheet as well as PDF format). Price bid in excel format should only be uploaded in 2nd cover. If bidder uploads price bid in PDF format along with technical bids/literature, the same will be</p>	<p>Price should be quoted in second cover on line in prescribed format as per Appendix-10 (Excel sheet as well as PDF format)</p>	<p>It is clarified that Price should be quoted in prescribed format in Excel sheet only as per Appendix-10 .</p> <p>However , it may be read as Excel Sheet only wherever it is mentioned in GTE.</p>

Handwritten mark

	<p>rejected. The quoted rates must be valid for a period of 180 days from the date of opening of tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected. In case the tenderer withdraws, modifies or change his offer during the validity period, bid is liable to be rejected and the EMD shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, condition etc. of their original tender.</p>		
<p>7.</p>	<p>Annexure-II Clause 14(i) at page No.20 of GTE</p> <p>Existing Clause:- Only the technical bid shall be opened online on the date of tender opening. Price bids of only those firms will be considered for opening whose offer would meet all tender requirements and TE specifications and has passed in all tender conditions and technical/physical (including field trial) evaluation if required.</p>	<p>-</p>	<p>It was decided during the meeting that while evaluation of tender, field trials/demonstration of the equipment will be carried out as per specifications. The bidder shall be ready for the field trials/demonstration in short notice after receiving of intimation.</p>

2. All other terms and conditions will remain unchanged. The bidders may submit their bids accordingly.


 (Rajendra Singh) DC(Proc)
For and on behalf of the President of India

copy to: -

1	<p>Indenter - (i) Director , CFSL, CBI Block No 4, CGO Complex, Lodi Road, New Delhi 110003.</p> <p>(ii) Director, CFSL, DFSS, Block No 9, 8th Floor, CGO Complex, Lodi Road, New Delhi 110003.</p> <p>(iii) Director, LNJN, NICFS, Outer Ring Road, Sector-03, Rohini, Delhi – 110085.</p>	For information with reference to this office of even No 06.12.2018.
2	Ministry of External Affairs, UOI New Delhi.	With requested to forward copy of corrigendum notice to the Ambassador / High commissioner of Embassies of those countries who are likely suppliers of subject item.
3	The Section Officer (IT), IT Cell Website room no. 10, MHA, North Block, New Delhi-110001.	With requested to get it incorporated in the official website of the Ministry of Home Affairs.