

GOVERNMENT OF INDIA
MINISTRY OF HOME AFFAIRS
(PROCUREMENT WING)
JAISALMER HOUSE, 26 MAN SINGH ROAD,
NEW DELHI-110 011

Dated 04.01.2019

CORRIGENDUM NO. 01
TO GLOBAL TENDER ENQUIRY (GTE)
No. D/21013/3274/27.09.2018/P-1 Dated 17.12.2018

REF: GLOBAL TENDER ENQUIRY (GTE) NO. D/21013/3274/27.09.2018/P-1 DATED 17.12.2018 FLOATED ON CPP PORTAL AND SCHEDULE TO BE OPENED ON 22.01.2019 (TECH BID) FOR (I) UV-VISIBLE SPECTROPHOTOMETER-QTY. 06 NOS (II) GAS CHROMATOGRAPHY-MASS SPECTROMETRY (GC-MS) SYSTEM-QTY. 06 NOS AND (III) GAS CHROMATOGRAPHY SYSTEM(GC) - QTY 06 OF EACH ITEM FOR CFSL - BHOPAL, CHANDIGARH, GUWAHATI, HYDERABAD, KOLKATA AND PUNE OF DFSS.

A pre-bid meeting under the Chairmanship of Director of DFSS, New Delhi was held on 26.12.2018 at 11.30 AM to address the points raised by prospective bidders in connection with Global Tender Enquiry (GTE) issued by MHA on 17.12.2018 floated on CPP portal for procurement of above items for CFSLs, of DFSS. After discussion in the pre-bid meeting, after approval of competent authority, following amendments are hereby issued against the subject GTE. These changes may be effected wherever it is defined in the text of the tender documents and it may be treated as amended accordingly:-

(Minutes of Pre-Bid Meeting)

S.No.	Ref. of Para in GTE	Clarification sought/ Suggestion submitted by Rep. of firms	Changes agreed not/Not agreed by MHA/Amendment in GTE
1.	Appendix-2, Sl. No. 7 at Page No. 34 of GTE: Existing Clause:- Absorbance(A),%Transmittance,% Reflectance and Concentration(C)"	M/s Toshvin Analytical Pvt. Ltd., Request this may be changed Concentration(C)/ Quantitation	Agreed For Existing clause To be read as :- Absorbance(A),%Transmittance, % Reflectance and Concentration(C)/Quantitation".
2.	Appendix-2, Sl. No. 10 at Page No. 34 of GTE: Existing Clause:- Stray light At 198 nm (for KCl) 0.5% T or better At 220 nm NAI(ASTM method) 0.0001 % T or better At 340 nm (for NaNo ₂) 0.0001% T or better	M/s Toshvin Analytical Pvt. Ltd., has requested that, Globally reputed vendors has 1% T at 198 nm & In our catalogue we do not mention ASTM. Hence, this may be change At 198 nm (for KCl)1% T or better At 220 nm NAI 0.0001 % T or better	Not agreed Tender Condition Persists.
3.	Appendix-2, Sl. No. 15 at Page No. 34 of GTE:	M/s Toshvin Analytical Pvt. Ltd.,	Agreed

R

3.	<p>Appendix-2, Sl. No. 15 at Page No. 34 of GTE:</p> <p>Existing Clause:- Data Station: The principal should supply the required branded (DELL/HP/IBM etc) computer system along with licensed operating system with 21" LED monitor and laser jet printer.</p>	<p>M/s Toshvin Analytical Pvt. Ltd.,</p> <p>It should be i7 process, Original Window 10 , 16 GB RAM, 1 TB HDD, 21 inch monitor, antivirus,</p> <p>As a apex govt. body it is always recommend to go for top end PC.</p>	<p>Agreed</p> <p>For Existing clause</p> <p>To be read as :-</p> <p>Data Station: Latest version of branded PC like DELL/HP/COMPAQ/LENOVO/IBM etc with i5 processor or higher version, minimum 4 GB RAM and 1 TB internal memory capacity or better, along with the license of original operating system and preloaded software with 21" LED monitor and LaserJet printer.</p>
4.	<p>Appendix-2, Sl. No. 17 (viii) at Page No. 35 of GTE:</p> <p>Existing Clause:- viii) Comprehensive and non-comprehensive AMC rate must be quoted for 5 years from the post warranty date payable on yearly basis. Preventive maintenance- 4 visit per year on quarterly basis and breakdown visits as and when required. The supplier should also certify that trained service engineers are available to attend the emergency visits within a downtime of 48 hours.</p>	<p>During the pre-meeting vendors has raised that point in the commercial part of GTE AMC rate have been mentioned, whereas in technical specification Comprehensive and non-comprehensive AMC rate have been mentioned. Requested to specify.</p>	<p>It is clarified that as per Annexure IV clause 9 at page no 31 of GTE it is AMC for 5 years after warranty period of 2 years. Hence, rate for AMC to be quoted as per BOQ after warranty period of 2 years.</p>
5.	<p>Appendix-2, Sl. No. 1 at Page No. 34 of GTE:</p> <p>Existing Clause:- OPTICS Double Beam UV Visible Spectrophotometer</p>	<p>M/s PerkinElmer</p> <p>has requested to change the Double Beam UV Visible spectrophotometer with Double monochromator.</p>	<p>Not agreed</p> <p>Tender Condition Persists.</p>

M

**(B) Name of the equipment: Gas Chromatography-Mass Spectrometer (GC-MS) System.
(Technical parameters)**

S.No	Ref. of Para in GTE	Clarification sought/ Suggestion submitted by Rep. of firms	Changes agreed not/Not agreed by MHA/Amendment in GTE
1.	<p>Appendix-2, Sl. No. 9 (Sensitivity/detection limit) at Page No. 37 of GTE:</p> <p>Existing Clause:-</p> <p>1 Pg std OFN at m/z with scanning from 50-300 amu with Helium as carrier gas should produce S/N 800:1 or higher (installation specification) in EI mode and equivalent or with any other compound.</p>	<p>M/s Toshvin Analytical Pvt. Ltd has request that not to further dilute the specification of S/N ratio. Further,</p> <p>& M/S Thermo Fisher Scientific India Pvt. Ltd., has requested that Instruments detection limit also alongwith signal to noise ratio. Suggested to mentioned Instruments detection limit shall be 5 fg or better for OFN</p>	<p>Not agreed</p> <p>Tender Condition Persists.</p>
2.	<p>Appendix-2, Sl. No. 15 at Page No. 37 of GTE:</p> <p>Existing Clause:-</p> <p>Mass Library (with license)</p> <p>a) NIST Mass spectral library</p>	<p>M/s Toshvin Analytical Pvt. Ltd</p> <p>This is the latest version and should be clearly mentioned in the tender specification NIST 2017, hence proposed requested for an amendment as under: Original Licensed NIST 2017 or above in CD ROM format.</p>	<p>Agreed</p> <p>For Existing clause</p> <p>To be read as :-</p> <p>Mass Library (with license) latest version</p> <p>a) NIST Mass spectral library</p>
3.	<p>Appendix-2, Sl. No. 17 at Page No. 37 of GTE:</p> <p>Existing Clause:-</p> <p>Data Station: Latest version of branded PC like DELL/HP/COMPAQ/LENOVO/IBM make along with the license of original operating system with 21" LED monitor and laser jet printer..</p>	<p>The representatives of firm M/s Toshvin Analytical Pvt. Ltd has requested that it should be i7 process, Original Window 10, 16 GB RAM, 1 TB HDD, 21 inch monitor, antivirus and as apex govt. body it is always recommend to go for top end PC.</p>	<p>Agreed</p> <p>For Existing clause</p> <p>To be read as :-</p> <p>Data Station: The principal should supply the required branded (DELL/HP/IBM etc. with i5 or higher processor, minimum 4 GB RAM and 1 TB internal memory or better. Computer system along with licensed operating system with 21" LED monitor and LaserJet printer. The original equipment supplier / manufacturer should supply the computer data station preloaded in the factory.</p>

Ru

4	<p>Appendix-2, Sl. No. 18 at Page No. 37 of GTE:</p> <p>Existing Clause:-</p> <p>ii) Comprehensive and non-comprehensive AMC rate must be quoted for 5 years from the post warranty date payable on yearly basis. Preventive maintenance- 4 visit per year on quarterly basis and breakdown visits as and when required. The supplier should also certify that trained service engineers are available to attend the emergency visits within a downtime of 48 hours</p>	<p>During the pre-meeting vendors has raised that point in the commercial part of GTE AMC rate have been mentioned, whereas in technical specification Comprehensive and non-comprehensive AMC rate have been mentioned. Requested to specify.</p>	<p>It is clarified that as per Annexure IV clause 9 at page no 31 of GTE it is AMC for 5 years after warranty period of 2 years. Hence, rate for AMC to be quoted as per BOQ after warranty period of 2 years.</p>
---	---	---	--

**(C) Name of the equipment: Gas Chromatography (GC) System.
(Technical parameters)**

S.No	Ref. of Para in GTE	Clarification sought/ Suggestion submitted by Rep. of firms	Changes agreed not/Not agreed by MHA/Amendment in GTE
1.	<p>Appendix-2, Sl. No. 5 at Page No. 39 of GTE:</p> <p>Existing Clause:- Flow controller for carrier and carrier and detector</p> <p>Automatic digital flow and pressure controller/EPC/EFC/IEC must be available</p>	<p>M/s Toshvin Analytical Pvt. Ltd proposed to request also add AFC</p> <p>Automatic digital flow and pressure controller/AFC/EPC/EFC /IEC must be available.</p>	<p>Agreed</p> <p>For Existing clause</p> <p>To be read as :-</p> <p>Flow controller for carrier and detector</p> <p>Automatic digital flow and pressure controller/EPC/EFC/IEC/AFC must be available</p>
2.	<p>Appendix-2, Sl. No. 13 at Page No. 39 of GTE:</p> <p>Existing Clause:- Data Station: The principal should supply the branded (DELL/HP/IBM etc) computer system along with license of operating system with 21" branded LED monitor along with laser jet printer.</p>	<p>The representative of firm M/s Toshvin Analytical Pvt. Ltd has requested that it should be i7 process, Original Window 10, 16 GB RAM, 1 TB HDD, 21 inch monitor, antivirus.</p>	<p>Agreed</p> <p>For Existing clause</p> <p>To be read as :-</p> <p>Data Station: The principal should supply the required branded (DELL/HP/IBM etc. with i5 or higher processor, minimum 4 GB RAM and 1 TB internal memory or better. Computer system along with licensed operating system with 21" LED monitor and laserjet printer. The original equipment supplier /</p>



			manufacturer should supply the computer data station preloaded in the factory.
3.	<p>Appendix-2, Sl. No. 9(Detector) at Page No. 39 of GTE:</p> <p>Existing Clause:- Flame Ionization Detector-02 Nos. The system should have two no.s of detectors (FID) mounted simultaneously. FID- Temp Max-4000°C Dynamic range 10⁷ Minimum detection limit: 2pgC/s</p>	<p>M/s PerkinElmer has requested that it may be amended :- FID Temp Max: 400°C Dynamic range : 10⁷ MDL : 3.10pgC/s</p>	<p>To be read as :-</p> <p>Flame Ionization Detector-02 Nos. The system should have two no.s of detectors (FID) mounted simultaneously. FID- Temp Max-400°C Dynamic range 10⁷ Minimum detection limit: 2pgC/s</p>
4.	<p>Appendix-2, Sl. No. 16 at Page No. 40 of GTE:</p> <p>Existing Clause:- i) Comprehensive and non-comprehensive AMC rate must be quoted for 5 years from the post warranty date. AMC includes Preventive maintenance-4 visits per year on quarterly basis and breakdown visit as & when required.</p>	<p>During the pre-meeting vendors has raised that point in the commercial part of GTE AMC rate have been mentioned, whereas in technical specification Comprehensive and non-comprehensive AMC rate have been mentioned. Requested to specify.</p>	<p>It is clarified that as per Annexure IV clause 9 at page no 31 of GTE it is AMC for 5 years after warranty period of 2 years. Hence, rate for AMC to be quoted as per BOQ after warranty period of 2 years.</p>

(Commercial Parameters of GTE)


S.No	Ref. of Para in GTE	Clarification sought/ Suggestion submitted by Rep. of firms	Changes agreed not/Not agreed by MHA/Amendment in GTE
1	<p>Annexure II Clause 1(D), at Page No. 9 of GTE:</p> <p>EMD: Hard copy of original EMD instrument should be deposited offline directly to Office of Joint Secretary(PM),(Procurement Wing), PM Division, MHA, Jaisalmer House, 26 Man Singh Road, New Delhi-110011 on or before 21.01.2019 at 03.30 hours and scanned copy of the same should be uploaded along with other tender documents failing which their offer will be summarily rejected.</p>	<p>M/s Thermo Fisher Scientific India Pvt. Ltd., has requested that EMD should be asked instrument wise so that interested vendor can submit EMDs for the produce they want to bid for.</p>	<p>Not agreed</p> <p>Tender Condition Persists.</p>

re

	<p>failing which their offer will be summarily rejected.</p>		
2	<p>Annexure II Clause 6.2 (c), at Page No. 14 of GTE: Evaluation Criteria:- a. The tenderer are required to bid for entire quantity of stores as per Bill of Quantity (BOQ) at Appendix-10 of GTE. The lowest bid will be decided upon the lowest price quoted by the bidder as on the date of opening of Price bid. The Lowest Price (L-1) for both foreign and indigenous Bidders will be decided upon the lowest price quoted by the particular Bidder (item-wise) as on date of opening of tender enquiry as per the Price Format and as explained below :</p>	<p>M/s Thermo Fisher Scientific India Pvt. Ltd has requested that there should be provision for quoting separate BOQ per instrument. Also BOQ should contain option to quote in foreign currency and local items in INR.</p>	<p>It was clarified during the pre-bid meeting that as per clause 6.2(i) L-1 status will be decided on the basis of item wise price quoted by the particular bidder inclusive of all applicable levies/duties/ taxes/ agency commission etc. BOQ has already been designed for L-1 status for separate item as quoted by the bidders. Hence, no change is required.</p>
3.	<p>Payment Terms : Annexure IV Clause 6A(i), at Page No. 30 of GTE: (i) 100% payment shall be paid on receipt of goods at destination, successful installation, commissioning , training and final acceptance of the goods by the consignee in terms of contract and submission of performance guarantee for 10% of the total contract value valid till warranty period</p>	<p>M/s Thermo Fisher Scientific India Pvt. Ltd has requested (i) to pay 80% advance against purchase order and 20% after installation for local vendors in INR. (ii) To foreign manufacturers : 80% payment of stores shall be paid against irrevocable letter of credit (LC) on presentation of required documents (as mentioned in tender documents) and submission of performance guarantee for 10% of contract value valid till warrant period and balance 20% on certification of successful installation, commissioning, training and acceptance of the goods by the consignee.</p>	<p>Amendment: For : Existing entries Read: (i) FOR IMPORTED GOODS: To Foreign seller: 80% net value of stores for delivery at destination will be paid on receipt of item in good and acceptable condition by the consignee at destination CIP/FOR and on submission of performance guarantee for 10% of total contract value valid till warranty period and balance 20% payment shall be paid on successful installation, commissioning, training and final acceptance of the goods by the consignee. To Indian Agent: 100 % payment of agency commission, if any, as quoted in price bid, shall be paid in Indian Rupees after successful installation, commissioning, training and final acceptance of the goods by the consignee in terms of contract.</p>

			<p>(ii) FOR DOMESTIC GOODS:</p> <p>To Indigenous seller: 80% net value of stores for delivery at destination will be paid on receipt of item in good and acceptable condition by the consignee at destination CIP/FOR and on production of required documents (as mentioned in tender documents) and submission of performance guarantee for 10% of total contract value valid till warranty period and balance 20% payment shall be paid on successful installation, commissioning, training and final acceptance of the goods by the consignee.</p> <p>It may be read as wherever it is mentioned in GTE.</p>
4.	Add new clause at Annexure-III (in General Conditions) after clause 24 at page No. 25	-	<p>Read</p> <p>Clause 25. At Annexure-III of GTE:-</p> <p>All the technical specifications of the equipment should either be available with the attached technical specifications sheet of the manufacture otherwise a certificate from the original equipment manufacturer must be attached.</p>

2. All other terms and conditions will remain unchanged. The bidders may submit their bids accordingly.


(Rajendra Singh) DC(Proc)
For and on behalf of the President of India

copy to: -

1	<p>Indenters – (1) The Coordinator, Central Forensic Science Laboratory, Gomantika Parisar, Jawahar Chowk, Bhopal, (M.P) – 462003.</p> <p>(2) Director, Central Forensic Science Laboratory, Dakshin Marg, Sector 36-A, Chandigarh- 160036.</p> <p>(3) Director-Cum-Coordinator, Central Forensic Science Laboratory, 16, Lachit Borphukan Path, Tetelia, Gotanagar, Guwahati – 781011.</p> <p>(4) Director, Central Forensic Science Laboratory, Ramanthapur, Hyderabad, Telangana -500013.</p> <p>(5) Director, Central Forensic Science Laboratory, 30, Gorachand Road, Kolkatta – 700014.</p> <p>(6) The Coordinator, Central Forensic Science Laboratory, 38/4, Krshna Complex, Kharadi Bypass, Kharadi, Pune – 411014.</p>	
2.	Director, Dte. Of Forensic Science Block No. 09, 8 th Floor, CGO Complex, Lodhi Road, New Delhi-110003.	For information with reference to this wing GTE of even No. dated 17.12.2018.
3	Ministry of External Affairs, UOI New Delhi.	With requested to forward copy of corrigendum notice to the Ambassador / High commissioner of Embassies of those countries who are likely suppliers of subject store.
4	The Section Officer (IT), IT Cell Website room no. 10, MHA, North Block, New Delhi-110001.	With requested to get it incorporated in the official website of the Ministry of Home Affairs.